

PCB Piezotronics Inc. claims proprietary rights in the information disclosed herein. Neither it nor any reproduction thereof will be disclosed to others without the written consent of PCB Piezotronics Inc.

REVISIONS

REV	DESCRIPTION	DIN
F	CORRECT UNIT CONVERSION	52254

18893

ELECTRICAL CONNECTOR
10-32 UNF - 2A

2 MODEL 060A14
CLAMP NUT 14mm HEX
ST STL (SUPPLIED)

M14x1.25 - 6g

MODEL 065A37
SEAL .025 [0.64] THICK
BRASS (SUPPLIED)

1.65 [41.9] MAX

.900 [22.86]

.300 [7.62]

Ø.435 [Ø11.05]

Ø.497 [Ø12.62]

MOUNTING HOLE PREPARATION

1 Ø.437±.001 [11.10±.025] THRU
LØ.500 [12.70] ▽.600 [15.24]
M14x1.25-6H ▽.450 [11.43]

.875 [22.23] 3

.002

A

⊥.005A

- 3 DIMENSIONS SHOWN ARE FOR .875 [22.23] THICK WALL. COUNTERBORE FOR THICKER WALLS
- 2 RECOMMENDED MOUNTING TORQUE ON A 14mm HEX: 5-6 FT LBS (7-8 NEWTON METERS)
- 1 SEAL SURFACE MUST BE FREE OF TOOL MARKS WITH A MINIMUM $\sqrt{63}$

UNLESS OTHERWISE SPECIFIED TOLERANCES ARE:

DRAWN		CHECKED		ENGINEER	
KSR	10/22/21	KSR	10/22/21	RF	10/22/21

PCB PIEZOTRONICS
AN AMPHENOL COMPANY

3425 WALDEN AVE. DEPEW, NY 14043
(716) 684-0001 E-MAIL: sales@pcb.com

INSTALLATION DRAWING
MODEL M106B
PRESSURE SENSOR

CODE IDENT. NO. 52681
DWG. NO. 18893

SCALE: 2X SHEET 1 OF 1